

PRICE BID FOR ANASTHESIA DEPARMENT EQUIPMENTS

EarnestMoneyRs...../-enclosed
NSC/FDR/CDRNo.

Security money 10% of the cost
Of the article.

Department-ANASTHESIA

| Sl.No. | Name of the article& Technical Specification | Cost of the article per unit |
|--------|--|------------------------------|
|--------|--|------------------------------|

4. Peripheral Nerve Stimulator

Technical specification

1. Should have a current range from 0.1 to 5mA.
2. Should have a resolution of 0.01mA
3. Should have a digital display for current.
4. Should have short stimulus pulse duration of 0.1ms.
5. Should have a frequency range from 0.1 to 99 Hz.
6. Should be able to monitor single twitch, Train of four, PTC and DBC.
7. Should have integrated electrode cable with lead.
8. Should be battery operated.
9. Should be supplied with 5 sets of insulated needles in varied length.
10. Should have safety certificate from a competent authority CE / FDA (US) STQC CB certificate, STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.

- 1-SHOUTABLEUPS/CVT
- 2-Warranty/CMCDATEDONINTALLATION5YEARS.

Signature of Tenderer
 Signature of Witness:
 Name:
 Full Address:

Name of the Tenderer:
 Address of Firm
 With stamp