

**PRICE BID FOR ORTHOPAEDIC DEPARTMENT EQUIPMENTS**

EarnestMoneyRs...../-enclosed  
NSC/FDR/CDRNo.

Security money 10% of the cost  
Of the article.

**Department-ORTHOPAEDIC**

Sl.No.	Name of the article& Technical Specification	Cost of the article per unit
--------	--	------------------------------

<b>HAND DRIVEN BATTERY OPERATED DRILL WITH FLEXIBLE REAMER AND K WIRE ATTACHMENT WITH POWER SAW - 2 SETS</b>		
<b>Specification :</b>		<b>QTY</b>
<b>Five Year Full Service Warranty (CMC)</b>		
<b>THE CANNULATED HANDPIECE</b>		<b>1</b>
THE DIAMETER OF CANNULATION SHOULD BE 4-5 mm		
MAXIMUM SPEED WITH ATTACHMENTS O TO BE BETWEEN 1000-1500 rpm		
HAND PIECE SHOULD BE COMPATABLE WITH RADIOLUCENT DRIVE		
INSTANT CHANGE BETWEEN CLOCK WISE AND COUNTER CLOCK WISE ROTATION SHOULD BE		
POSSIBLE THE SYSTEM SHOULD BE FULLY AUTOCLAVABLE		
ALL ATTACHMENT CAN BE FITTED ON SINGLE HAND PIECE		
RELIABLE PROTECTION OF SOFT TISSUE WITH INTEGRATED ASCILLATION MORE		
<b>POWER MODULE</b>		<b>1</b>
MODE SLECTOR SWITCH TO BE SELECT DRILLING/REAMING, SAW, OSCILLATING DRILL MODE		
THERE SHOULD BE A DISPLAY INDICATING THE BATTERY CAPACITY STATUS.		
SHOULD HAVE A BUTTON TO DIAGNOSE ERRORS IN THE SYSTEM.		
<b>LID FOR HAND PIECE</b>		<b>1</b>
FULLY AUTOCLAVABLE		
FOR HOLDING AND PREVENTING FALL AO POWER MODULE FROM THE HAND PIECE		
MODE SELECTOR SWITCH TO SELECT DRILLING/REAMING, SAW, OSCILLATING DRILL MODE		
<b>BATTERIES</b>		<b>1</b>
MAXIMUM CHARGING TIME SHOULD BE BETWEEN 20-40 MIN		
THE BATTERIES SHOULD BE SAFE AND EASY TO HANDLE IN THE OPERATING THEATRE		
THE BATTERIES SHOULD BE LITHIUM ION BATTERY WITH A MINIMUM LIFE OF 900-1200 CYCLES		
<b>STRILE COVER</b>		<b>1</b>
FULLY AUTO CLAVABLE		
<b>UNIVERSAL BATTERY CHARGER II</b>		<b>1</b>
SHOULD BE CAPABLE OF CHARGING NiCD, NiMh AND LITHIUM ION BATTERIES		
<b>QUICK COUPLING ATTACHMENT</b>		
<b>JACOB'S CHUCK ATTACHMENT</b>		<b>1</b>
<b>QUICK COUPLING FOR MODULLARY REAMING</b>		<b>1</b>
THE DIAMETER OF CANNULATION SHOULD BE 4-5 mm		
<b>SCREW ATTACHMENT WITH QUICK COUPLING</b>		<b>1</b>

<b>QUICK COUPLING FOR DRILLING</b>	<b>1</b>
<b>SAGITAL SAW ATTACHMENT</b>	<b>1</b>
SAW BLADE FOR TKR SURGERY	
SAW BLADE FOR GENERAL TRAUMATOLOGY	
THICKNESS 0.4 mm TO 1.47 mm (SET OF 5 BLADES)	
<b>FLEXIBLE REAMERS &amp; SAW WITH GUIDE WIRE EXCHANGER</b>	
Each Unit Should Contain:	
Pistol Grip Saggital Saw (With Set Of Five Blades)	
Flexible Reamer Fixed Head Shaft One Each	
7mm	
8mm	
9mm	
10mm	
11mm	
12mm	
13mm	
14mm	
15mm	
Flexible Reamer Should Be Kept In Alluminum Box	
<b>GUIDE WIRE EXCHANGER SLEEVE</b>	<b>1</b>
If Demonstration Required Than Should Ready To Live Demo	
<b>Company Should Be ISO 13485-2003/ FDA CERTIFIED</b>	




1-SHOUTABLEUPS/CVT

2-Warranty/CMDATEDONINTALLATION5YEARS.

Signature of Tenderer

Signature of Witness:

Name:

Full Address:

Name of the Tenderer:

Address of Firm

With stamp