

PRICE BID FOR OBSTETRICS & GYNAECOLOGY DEPARTMENT EQUIPMENTS


EarnestMoneyRs...../-enclosed
NSC/FDR/CDRNo.

Security money 10% of the cost
Of the article.


Department- OBSTETRICS & GYNAECOLOGY

Sl.No.	Name of the article& Technical Specification	Cost of the article per unit
	<ul style="list-style-type: none"> • Anti Hypoxic Rotameter- its pneumatic device Ensures no nitrous oxide flow, if oxygen flow is lower than 0.5 ltr/min approx. Minimum 25% (±5%) oxygen in surgery gas mixture. • Ratio control 1:3 mixed ratio between O₂ & N₂O gas flow. • Oxygen driven N₂O supply. • Oxy failure warning device (OFWD). • Higher efficiency forged regulators & yokes with SS fittings. • High precision pressure relief cum non return valve. • Vaporizers Gold man Halothane. • Change over mechanism with oxygen flush (open & close circuit). • Gas Inlet Port O₂ & N₂O Each 3 nos. Pin-indexed gas cylinder Connections for O₂ & N₂O each 2nos & diameter index gas specific Central Pipe Line Connection each 1No. for O₂, N₂O. • Standard Magills circuit, all rubber antistatic tested. • Trolley made of M.S Powder coated, Stainless steel table Top Tray , Modified monitoring accessories tray, drawer top stainless steel tray. • 5" dia antistatic breaking castor wheel. • Oxygen outlet for driving ventilator (Double). • Higher visibility & clarity Large diameter (63mm) O₂ & N₂O gauge each two nos. • O₂ & N₂O operating pressure & pipeline pressure (4.22Kgf/cm²) display gauge each one no. • 10 liter W/C cylinder carrier facility. • Provision for Tec Vaporizer, Circle Absorber & Ventilator, Ether Unit. (Optional at extra cost). • Size 1490 mm X 910 mm X 750 mm (H x W x D). 	

Tec	Meditec Pisces Tec Vaporizer Selecta Tec compatible with interlock generates accurate concentration under variable temp. & flow rate, low body weight service. Incorporate Graduation 0.2%-5%
Vaporizer	
Meditec brand	Drug Capacity – Max 225ml. ± 25ml. Drug Filling processes Screw cap filling / Pour fill Flow Range – Operating flow range 0.2 to 15lits/min Operating Temp. Range 15° to 35°C. Dimension: 114W x 220D x 235H mm.


 Professor & Head of the Department of
 OBSTETRICS & GYNACOLOGY
 M.J.B. Medical College, RAIPUR


 डॉ० सुशीला खर्कवाल
 एम० डी०
 आचार्य, प्रसूति एवं स्त्री रोग विभाग
 म० ल० बा० मेडिकल कालेज इण्डिया


 डॉ० विद्या वीरवार
 M.S., FICOG, MCH
 आचार्य, स्त्री रोग विभाग
 म० ल० बा० मेडिकल कालेज इण्डिया

1-SHOUTABLEUPS/CVT

2-Warranty/CMC DATED ON INTALLATION 5 YEARS.

Signature of Tenderer

Signature of Witness:

Name:

Full Address:

Name of the Tenderer:

Address of Firm

With stamp