

PRICE BID FOR SURGERY DEPARTMENT EQUIPMENTS

Earnest Money Rs..... /- enclosed
NSC/FDR/CDR No.

Security money 10% of the cost as
of the article.

Department - Surgery

Sl.No.	Name of the article & Technical Specific	Cost of the article per unit
--------	--	------------------------------


Laparoscopic Set Specifications – High Definition Endovision System

Required state-of-the-art High Definition Endovision System including the following components:

1. **High Definition Camera System**
 - Pure 1090p system - 1920 x 1080 pixels native resolution (lower resolution will not be acceptable).
 - Three Chip Camera Head with Progressive Scanning.
2. **High Definition Monitor**
 - Medical Grade Flat Screen LED Monitor for use with HD System.
 - Minimum 25 inches (lower size will not be acceptable).
3. **LED light source**
 - Illumination=300 watt
 - Minimum 5000 hrs life or more
 - Intensity control
 - Light weight.
4. **High Capacity Endoflator**
 - At Least 30-40Litres Capacity (lower capacity will not be acceptable).
 - Capable of heating gas option.
 - Must have digital display for all preset and real time values.
 - Must have automatic alarm and pressure control system in case of over pressure.
5. **Telescope**
 - 10mm, 30 Degrees
 - 10mm, 0 Degree
 - 5mm, 30 Degree
 - High quality – for use with HD system
6. **Lap Instruments**
 - a. Trocar Cannula 10mm - 1
 - b. Trocar Cannula 5mm - 2
 - c. Atraumatic Grasper 5mm - 1
 - d. Traumatic 5mm - 1
 - e. Clip applicator 10 mm for LT300 and LT 400 Clip - 1
7. **Trolley for above equipment and monitor arm.**

Note:

1. All the items (including Light Source and Light Cable) **must be from a single manufacturer of international repute.** No local accessories would be acceptable. No "Economy Class", "E-Class" or "E-Line" model shall be accepted.
2. A list of at least five government medical colleges (with details of department and user) of very high repute should be included. Only the biggest institutes of national repute such as All India Institute of Medical Sciences, Sanjay Gandhi Post Graduate Institute, King George Medical College, PGI Chandigarh, CMC Vellore etc. shall be considered as references.
3. USFDA approved certificate to be attached for item number 1 to 6.
4. HD (1080 regulation) video and image recording system of good quality

1-SHOUTABLEUPS/CVT

2-Warranty/CMCDATEDONINTALLATION5YEARS.

Signature of Tenderer

Signature of Witness:

Name:

Full address:

Name of the Tenderer:

Address of Firm

With stamp