

PRICE BID FOR EYE DEPARTMENT EQUIPMENTS

Earnest Money Rs..... /- enclosed
as NSC/FDR/CDR No.

Security money 10% of the cost
of the article.

Department - EYE

Sl.No.	Name of the article & Technical Specific	Cost of the article per unit
--------	--	------------------------------

KERATOMETER**SPECIFICATION-**


1. Should be Achromatic lenses.
2. Positive fixation permits rapid measurement of central corneal thickness.
3. Two way adjustable chin and head rests.
4. One position instrument and measures both meridians without changing the optical system.
5. Coincidence focusing system with positive measurement.
6. Simple vertical adjustable head to fit patient.
7. Should be Dual eye-level sighting system, facilitates horizontal alignment.


MEASUREMENT RANGES-


1	Corneal power	52- 36 D Diopter in 0.25 D step
2	Corneal astigmatism axis	0° - 180°
3	Corneal radius	6.4 - 9.4 mm in 0.05 mm step



ELECTRICAL SPECIFICATIONS-

1	Should be Light source	LED Bulb
2	Power supply Strictly	~110-230 V, 50-60 Hz


 ४० अतेशु कुमर
 (२० लो मिये (जेमेलर))


 डॉ० मो० अरशद
 (अलिखटेर जेमेलर)


 डॉ० के के मितुशिय
 (जेमेलर)

1- SHOUTABLE UPS/CVT

2- Warranty/CMC DATED ON INTALLATION 5 YEARS.

Signature of Tenderer

Signature of Witness:

Name:

Full address:

Name of the Tenderer:

Address of Firm

With stamp